



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

1 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.01 Lead / Lag to UPF, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Calibrator by Direct / Comparison Method	0.04 Wh to 32 kWh	0.0047 % to 0.4 %
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.01 Lead / Lag to UPF, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 Wh to 38.4 kWh	0.0047 % to 1.6 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 W to 32 kW	0.0047 % to 0.4 %
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 W to 38.4 kW	0.0047 % to 1.6 %
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VAh to 32 kVAh	0.0047 % to 0.03 %
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	4 kVAh to 38.4 kVAh	0.0047 % to 0.0162 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 2 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VA to 32 kVA	0.0047 % to 0.035 %
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 VA to 38.4 kVA	0.0047 % to 0.0162 %
9	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VARh to 32 kVARh	0.03 % to 0.4 %
10	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 VARh to 38.4 kVARh	0.0047 % to 1.6 %
11	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VAR to 32 kVAR	0.005 % to 0.4 %
12	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 VAR to 38.4 kVAR	0.0047 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

3 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
13	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.1 Lead / Lag to UPF, 40 V to 320 V, 10 mA to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 Wh to 115.2 kWh	0.008 % to 0.16 %
14	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 10 mA, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 W to 9.6 W	0.008 % to 0.8 %
15	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 10 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.012 W to 115.2 kW	0.008 % to 1.6 %
16	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 10 mA to 120 A, 0.1 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VARh to 115.2 kVARh	0.008 % to 0.16 %
17	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 10 mA, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.19 VAR to 9.6 VAR	0.008 % to 0.8 %
18	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 10 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.012 VAR to 115.2 kVAR	0.008 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

4 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
19	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.01 Lead / Lag to UPF, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 Wh to 96 kWh	0.0047 % to 0.4 %
20	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 W to 96 kW	0.031 % to 0.4 %
21	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VA to 96 kVA	0.0047 % to 0.031 %
22	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VARh to 96 kVARh	0.0047 % to 0.4 %
23	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VAR to 96 kVAR	0.0047 % to 0.4 %
24	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Active Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 Wh to 115.2 kWh	0.0047 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

5 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
25	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 W to 115.2 kW	0.0047 % to 1.6 %
26	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VAh to 96 kVAh	0.0047 % to 0.031 %
27	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	12 kVAh to 115.2 kVAh	0.0047 % to 0.0162 %
28	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 VA to 115.2 kVA	0.0047 % to 0.0162 %
29	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 VARh to 115.2 kVARh	0.0047 % to 1.6 %
30	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 VAR to 115.2 kVAR	0.0047 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

6 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
31	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 10 Hz to 1 kHz	Using 8½ Digit Multimeter with MPC by Comparison Method	30 µA to 100 mA	0.11 % to 0.053 %
32	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 10 Hz to 1 kHz	Using 8½ Digit Multimeter by Direct Method	30 µA to 100 mA	0.11 % to 0.053 %
33	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 45 Hz to 65 Hz	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.001 A to 100 A	0.045 % to 0.004 %
34	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 45 Hz to 65 Hz	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	100 A to 120 A	0.004 % to 0.01 %
35	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 1 kHz	Using 8½ Digit Multimeter by Direct Method	100 mA to 1 A	0.053 % to 0.094 %
36	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 1 kHz	Using 8½ Digit Multimeter with MPC by Comparison Method	100 mA to 1 A	0.053 % to 0.094 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 7 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
37	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 5 kHz	Using 8½ Digit Multimeter by Direct Method	1 A to 20 A	0.094 % to 0.12 %
38	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 5 kHz	Using 8½ Digit Multimeter with MPC by Comparison Method	1 A to 20 A	0.094 % to 0.12 %
39	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC High Current @ 50 Hz	Using High Current to Low Current Conversion, STD Current Transformer & Low Current Measured at STD Meter with High Current Source by Direct / Comparison Method	100 A to 6000 A	0.065 % to 0.26 %
40	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC High Voltage @ 50 Hz	Using High Voltage Divider with kV Meter with HV Source by Direct / Comparison Method	1 kV to 100 kV	1.46 % to 1.47 %
41	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter by Direct Method	1 mV to 100 mV	0.5 % to 0.018 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 8 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
42	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter with MPC by Comparison Method	1 mV to 100 mV	0.54 % to 0.019 %
43	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter by Direct Method	100 mV to 1000 V	0.018 % to 0.02 %
44	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter with MPC by Comparison Method	100 mV to 1000 V	0.019 % to 0.02 %
45	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 45 Hz to 65 Hz	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	30 V to 480 V	0.004 % to 0.002 %
46	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance @ 1 kHz	Using Precision LCR Meter with Decade Capacitance Box by Direct Method / Substitution Method	100 pF to 100 µF	0.1 % to 0.14 %
47	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance @ 100 Hz	Using Precision LCR Meter with MPC by Direct Method / Substitution Method	1 mF to 100 mF	0.14 % to 0.19 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

9 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
48	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	Harmonics @ (50 Hz, 40 V to 240 V, 0.05 A to 10 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	1st order to 39th order	0.5 %
49	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	Inductance @ 1 kHz	Using Precision LCR Meter with Decade Inductance Box by Direct Method / Substitution Method	100 µH to 10 H	0.1 % to 0.11 %
50	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	Power Factor @ (50 Hz, 40 V to 480 V, 0.01 A to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	(-) 1 PF to 1 PF	0.00001 PF
51	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	Resistance @ 1 kHz	Using Precision LCR Meter with Decade Resistance Box by Direct Method / Substitution Method	1 ohm to 10 kohm	0.1 % to 0.08 %
52	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Active Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mWh to 36 kWh	0.071 % to 1.61 %
53	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Active Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mW to 36 kW	0.071 % to 1.61 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 10 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
54	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Reactive Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mVArh to 36 kVArh	0.071 % to 1.61 %
55	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Reactive Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mVAr to 36 kVAr	0.071 % to 1.61 %
56	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 3 Wire, AC Active Energy @ (50 Hz, 63.5 V, 1 mA to 10 A, 0.5 Lag / 0.8 Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	95.25 mWh to 1.905 kWh	0.071 % to 0.061 %
57	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 3 Wire, AC Reactive Energy @ (50 Hz, 63.5 V, 1 mA to 10 A, 0.5 Lag / 0.8 Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	95.25 mVArh to 1.905 kVArh	0.071 % to 0.061 %
58	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Active Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mWh to 108 kWh	0.071 % to 1.61 %
59	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Active Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mW to 108 kW	0.071 % to 1.61 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 11 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
60	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Reactive Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mVArh to 108 kVArh	0.071 % to 1.61 %
61	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Reactive Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mVAr to 108 kVAr	0.071 % to 1.61 %
62	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 45 Hz to 65 Hz	Using 3Ø Power Calibrator by Direct Method	1 mA to 120 A	0.034 % to 0.025 %
63	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz	Using Multi Product Calibrator with 50 Turns Current Coil by Direct Method	20 A to 1000 A	0.5 % to 0.55 %
64	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multi Product Calibrator by Direct Method	1 mA to 100 mA	0.14 % to 0.087 %
65	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multi Product Calibrator by Direct Method	100 mA to 20 A	0.087 % to 0.21 %
66	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multi Product Calibrator by Direct Method	30 µA to 1 mA	0.58 % to 0.14 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 12 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
67	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Inductance @ 1 kHz	Using Decade Inductance Box by Direct Method	100 μ H to 1000 mH	0.6 % to 0.18 %
68	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	1 mV to 3 mV	1.7 % to 0.4 %
69	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	1 V to 1000 V	0.04 % to 0.063 %
70	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	100 mV to 1 V	0.06 % to 0.04 %
71	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	3 mV to 100 mV	0.4 % to 0.06 %
72	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	Using Decade Capacitance Box by Direct Method	1 μ F to 100 μ F	1.8 % to 0.5 %
73	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	Using Decade Capacitance Box by Direct Method	100 pF to 1 μ F	0.7 % to 1.8 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 13 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
74	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	1 mohm to 100 mohm	0.056 % to 0.026 %
75	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	10 µohm to 50 µohm	0.087 % to 0.057 %
76	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	100 mohm to 1 ohm	0.011 % to 0.021 %
77	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	50 µohm to 1 mohm	0.057 % to 0.056 %
78	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter by Direct Method	1 µA to 100 mA	0.05 % to 0.005 %
79	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter with MPC by Comparison Method	1 µA to 100 mA	0.053 % to 0.005 %
80	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter by Direct Method	1 A to 20 A	0.022 % to 0.047 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 14 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
81	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter with MPC by Comparison Method	1 A to 20 A	0.022 % to 0.057 %
82	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter by Direct Method	100 mA to 1 A	0.005 % to 0.022 %
83	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter with MPC by Comparison Method	100 mA to 1 A	0.005 % to 0.022 %
84	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC High Voltage	Using High Voltage Divider with KV Meter with HV Source by Comparison Method	1 kV to 100 kV	2.21 % to 1.53 %
85	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter by Direct Method	0.1 mV to 1 mV	0.12 % to 0.014 %
86	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter with MPC by Comparison Method	0.1 mV to 1 mV	0.13 % to 0.016 %
87	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter by Direct Method	1 mV to 10 V	0.014 % to 0.0005 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 15 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
88	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter with MPC by Comparison Method	1 mV to 10 V	0.016 % to 0.0005 %
89	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter by Direct Method	10 V to 1000 V	0.0005 % to 0.0007 %
90	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter with MPC by Comparison Method	10 V to 1000 V	0.0005 % to 0.0007 %
91	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	1 ohm to 100 kohm	0.0027 % to 0.001 %
92	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	1 ohm to 100 kohm	0.0027 % to 0.001 %
93	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	10 Mohm to 100 Mohm	0.004 % to 0.017 %
94	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	10 Mohm to 100 Mohm	0.0025 % to 0.017 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 16 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
95	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	100 kohm to 10 Mohm	0.001 % to 0.004 %
96	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	100 Mohm to 1000 Mohm	0.017 % to 0.1 %
97	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	100 kohm to 10 Mohm	0.001 % to 0.0025 %
98	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	100 Mohm to 1000 Mohm	0.017 % to 0.1 %
99	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter by Direct Method	1 mohm to 100 mohm	0.055 % to 0.011 %
100	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with Resistance Box by Substitution Method	1 mohm to 100 mohm	5.8 % to 0.034 %
101	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	1 ohm to 100 kohm	0.021 % to 0.001 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 17 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
102	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter by Direct Method	1 ohm to 100 kohm	0.021 % to 0.001 %
103	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with Resistance Box by Substitution Method	10 µohm to 1 mohm	5 % to 5.8 %
104	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter by Direct Method	100 mohm to 1 ohm	0.011 % to 0.021 %
105	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	100 mohm to 1 ohm	0.034 % to 0.021 %
106	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator by Direct Method	1 µA to 10 µA	2.34 % to 0.35 %
107	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator by Direct Method	10 µA to 100 mA	0.35 % to 0.016 %
108	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator by Direct Method	100 mA to 20 A	0.016 % to 0.14 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 18 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
109	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator with 50 Turns Current Coil by Direct Method	20 A to 1000 A	0.4 %
110	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Decade Megaohmbox Standard by Direct Method	100 kohm to 1000 Mohm	0.1 % to 0.21 %
111	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance @ 1 kV	Using High Resistance Standard Box by Direct Method	10 Tohm	5.4 %
112	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance @ Up to 5 kV	Using Decade Mega Ohm Box Standard by Direct Method	10 Gohm to 1 Tohm	2.6 % to 4.3 %
113	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multi Product Calibrator by Direct Method	1 mV to 100 mV	0.355 % to 0.01 %
114	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multi Product Calibrator by Direct Method	100 mV to 1000 V	0.01 % to 0.008 %
115	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 2 Wire	Using Multi Product Calibrator by Direct Method	1 ohm to 10 kohm	0.07 % to 0.01 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 19 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
116	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 2 Wire	Using Multi Product Calibrator by Direct Method	10 Mohm to 1000 Mohm	0.07 % to 1.74 %
117	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 2 Wire	Using Multi Product Calibrator by Direct Method	10 kohm to 10 Mohm	0.01 % to 0.07 %
118	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 4 Wire	Using Multi Product Calibrator by Direct Method	1 mohm to 1 ohm	6 % to 0.072 %
119	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 4 Wire	Using Multi Product Calibrator by Direct Method	1 ohm to 10 kohm	0.072 % to 0.01 %
120	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 4 Wire	Using Low Resistance Standard by Direct Method	10 µohm to 1 ohm	0.64 % to 0.063 %
121	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge CT Mode - Phase Error (Min) Secondary Input - 1 A & 5 A	Using CT - PT Bridge by Comparison Method	120 % to 1 %	0.7 minute to 1.2 minute
122	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge CT Mode - Ratio Error (%) Secondary Input - 1 A & 5 A	Using CT - PT Bridge by Comparison Method	120 % to 1 %	0.013 % to 0.024 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 20 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
123	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge PT Mode - Phase Error (Min) Secondary Input - 63.5 V & 110 V	Using CT - PT Bridge by Comparison Method	120 % to 80 %	0.8 minute to 0.9 minute
124	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge PT Mode - Ratio Error (%) Secondary Input - 63.5 V & 110 V	Using CT - PT Bridge by Comparison Method	120 % to 80 %	0.015 % to 0.016 %
125	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Phase Error (1 A to 3200 A with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	1.18 minute to 2.3 minute
126	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Phase Error (3200 A to 6000 A Primary with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	1.18 minute to 2.3 minute
127	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Ratio Error (1 A to 3200 A with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	0.025 % to 0.04 %
128	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Ratio Error (3200 A to 6000 A Primary with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	0.025 % to 0.04 %
129	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer Burden Box	Using Power Network Analyzer with Source by Direct Method	1 VA to 30 VA	0.07 % to 0.06 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 21 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
130	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Potential Transformer / Capacitor Voltage Transformer - Phase Error (220 V / 110 V to 33 kV / 110 V with Secondary - 63.5 V & 110 V)	Using Standard PT & Comparator with Source Primary Injection by Comparison Method	120 % to 80 %	2.18 % to 2.3 %
131	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Potential Transformer / Capacitor Voltage Transformer - Ratio Error (220 V / 110 V to 33 kV / 110 V with Secondary - 63.5 V & 110 V)	Using Standard PT & Comparator with Source Primary Injection by Comparison Method	120 % to 80 %	0.07 %
132	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Potential Transformer Burden Box	Using Power Network Analyzer with Source by Direct Method	1.25 VA to 200 VA	0.07 % to 0.06 %
133	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Turn Ratio	Using 6½ Digit DMM & 8½ Digit Multimeter with Turn Ratio Standard by V/V Method	0.8 Turn to 2000 Turn	0.09 % to 0.23 %
134	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Conductivity Meter (1 µS to 1000 mS)	Using Decade Resistance Box by Simulation Method	1 ohm to 1 Mohm	0.68 % to 0.6 %
135	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	pH Meter (0 pH to 14 pH)	Using Multi Product Calibrator by Simulation Method	(-) 420 mV to 420 mV	0.3 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 22 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
136	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT 100)	Using 8½ Digit Multimeter with MPC by Substitution Method	(-) 200 °C to 850 °C	0.02 °C
137	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT 100, PT 1000)	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 850 °C	0.018 °C
138	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT 1000)	Using 8½ Digit Multimeter with MPC by Substitution Method	(-) 200 °C to 630 °C	0.02 °C
139	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple J - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 210 °C to 1200 °C	0.05 °C
140	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple B - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	600 °C to 1820 °C	0.124 °C
141	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple B - Type	Using 8½ Digit Multimeter by Direct Method	600 °C to 1820 °C	0.12 °C
142	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple E - Type	Using 8½ Digit Multimeter by Direct Method	(-) 250 °C to 1000 °C	0.08 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 23 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
143	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple E - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 250 °C to 1000 °C	0.09 °C
144	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple J - Type	Using 8½ Digit Multimeter by Direct Method	(-) 210 °C to 1200 °C	0.046 °C
145	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple K - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 1372 °C	0.051 °C
146	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple K - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 1372 °C	0.055 °C
147	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple L - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 900 °C	0.037 °C
148	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple L - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 900 °C	0.039 °C
149	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple N - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 1300 °C	0.073 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

24 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
150	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple N - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 1300 °C	0.079 °C
151	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple R - Type	Using 8½ Digit Multimeter by Direct Method	50 °C to 1767 °C	0.066 °C
152	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple R - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	50 °C to 1767 °C	0.08 °C
153	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple S - Type	Using 8½ Digit Multimeter by Direct Method	50 °C to 1767 °C	0.075 °C
154	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple S - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	50 °C to 1767 °C	0.09 °C
155	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple T - Type	Using 8½ Digit Multimeter by Direct Method	(-) 250 °C to 400 °C	0.114 °C
156	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple T - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 250 °C to 400 °C	0.124 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

25 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
157	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple U - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 600 °C	0.045 °C
158	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple U - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 600 °C	0.05 °C
159	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD (PT 100)	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 800 °C	0.03 °C
160	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD (PT 1000)	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 630 °C	0.03 °C
161	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple B - Type	Using Multi Product Calibrator by Direct Method	600 °C to 1820 °C	0.22 °C
162	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple E - Type	Using Multi Product Calibrator by Direct Method	(-) 250 °C to 1000 °C	0.23 °C
163	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple J - Type	Using Multi Product Calibrator by Direct Method	(-) 210 °C to 1200 °C	0.15 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 26 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
164	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple K - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 1372 °C	0.141 °C
165	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple L - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 900 °C	0.243 °C
166	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple N - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 1300 °C	0.19 °C
167	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple R - Type	Using Multi Product Calibrator by Direct Method	50 °C to 1767 °C	0.265 °C
168	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple S - Type	Using Multi Product Calibrator by Direct Method	50 °C to 1767 °C	0.265 °C
169	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple T - Type	Using Multi Product Calibrator by Direct Method	(-) 250 °C to 400 °C	0.343 °C
170	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple U - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 600 °C	0.29 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

27 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
171	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter by Direct Method	1 Hz to 50 Hz	0.0002 % to 0.00013 %
172	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter with MPC by Comparison Method	1 Hz to 50 Hz	0.0007 % to 0.00015 %
173	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter by Direct Method	100 kHz to 2 MHz	0.0001 % to 0.00015 %
174	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter with MPC by Comparison Method	100 kHz to 2 MHz	0.00011 % to 0.00013 %
175	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter by Direct Method	50 Hz to 100 kHz	0.00013 % to 0.0001 %
176	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter with MPC by Comparison Method	50 Hz to 100 kHz	0.00015 % to 0.00011 %
177	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time	Using Time Interval Meter by Comparison Method	1 s to 9900 s	0.12 s to 0.53 s



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 28 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
178	ELECTRO- TECHNICAL- TIME & FREQUENCY (Measure)	Time	Using Time Interval Meter by Comparison Method	36000 s to 86400 s	2.07 s to 7.08 s
179	ELECTRO- TECHNICAL- TIME & FREQUENCY (Measure)	Time	Using Time Interval Meter by Comparison Method	9900 s to 36000 s	0.53 s to 2.07 s
180	ELECTRO- TECHNICAL- TIME & FREQUENCY (Source)	Frequency	Using Multi Product Calibrator by Direct Method	1 Hz to 50 Hz	0.59 % to 0.012 %
181	ELECTRO- TECHNICAL- TIME & FREQUENCY (Source)	Frequency	Using Multi Product Calibrator by Direct Method	50 Hz to 2 MHz	0.012 % to 0.03 %
182	MECHANICAL- ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 100 RPM to 5000 RPM	3.1 RPM
183	MECHANICAL- ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 5000 RPM to 60000 RPM	16.02 RPM
184	MECHANICAL- ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 60000 RPM to 99950 RPM	20.2 RPM
185	MECHANICAL- ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	6 RPM to 100 RPM	1 RPM



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 29 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
186	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Sensor with Indicator, Rotation Meter with Indicator (Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 1000 RPM to 11000 RPM	4.01 RPM
187	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Sensor with Indicator, Rotation Meter with Indicator (Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	6 RPM to 1000 RPM	1.01 RPM
188	MECHANICAL-ACOUSTICS	Sound Level Meter, dB Meter, Acoustic Meter, Noise Meter @ 1 kHz	Using Sound Level Calibrator by Comparison Method	94 dB & 114 dB	0.51 dB
189	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Angle Protractor, Inclinator, Digital Level and Protractor, Slop Meter, Angle Meter (L.C.: 0.01° and Coarser) - All Quadrant's	Using Angle Gauge Block Set by Comparison Method	0° - 90° - 0°	3 minute of arc
190	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Angle Template, Centre Gauge, Chamfer Gauge, Cross Hatch Cutter, Drill Angle Gauge, Drill Point Gauge - Angle (All Quadrant's)	Using Profile Projector by Comparison Method	0° to 180°	6 minute of arc
191	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bevel Protractor, Universal Protractor, Angle Protractor, Digital Protractor (L.C.: 1 minute and Coarser) - All Quadrant's	Using Angle Gauge Block Set by Comparison Method	0°-90°-0°	3 minute of arc



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 30 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
192	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bore Gauge with Dial Indicator - Transmission Error (L.C.: 0.001 mm)	Using Dial Calibration Tester by Comparison Method	0 to 1 mm	4 µm
193	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauges, Slip Gauge Accessories by Comparison Method	0 to 1000 mm	17 µm
194	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Caliper Checker by Comparison Method	0 to 150 mm	10 µm
195	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Gauge Block Set and Long Slip Gauges, Slip Gauge Accessories by Comparison Method	0 to 150 mm	9.1 µm
196	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Caliper Checker by Comparison Method	0 to 300 mm	11.5 µm
197	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauges, Slip Gauge Accessories by Comparison Method	0 to 600 mm	13.1 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

31 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
198	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Caliper Checker by Comparison Method	0 to 600 mm	14 µm
199	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper - Dial / Digital / Analog / Vernier (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauges, Slip Gauge Accessories by Comparison Method	0 to 300 mm	10 µm
200	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Coating Thickness Gauge (L.C.: 0.0001 mm)	Using Thickness Foils by Comparison Method	10 µm to 100 µm	1.5 µm
201	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Coating Thickness Gauge (L.C.: 0.001 mm)	Using Thickness Foils by Comparison Method	10 µm to 1999 µm	2.5 µm
202	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Combination Set (L.C.: 1°) - All Quadrant's	Using Angle Gauge Block Set by Comparison Method	0° to 180°	35 minute of arc
203	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Comparator Stand - Flatness	Using Digital Plunger Dial Gauge (L.C. 0.1µm) by Comparison Method	Up to 300 mm X 300 mm	3.2 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 32 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
204	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Cross Hatch Cutter - Pitch	Using Profile Projector by Comparison Method	0.05 mm to 3 mm	12.3 µm
205	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Degree Protractor (L.C.: 0.01°)	Using Angle Gauge Block Set by Comparison Method	0° - 90° - 0°	3 minute of arc
206	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Gauge (Dial Type), Dial Depth Gauge (L.C.: 0.001 mm)	Using Gauge Block Set, Long Slip Gauges, Surface Plate by Comparison Method	0 to 300 mm	6.5 µm
207	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set, Long Slip Gauges, Surface Plate by Comparison Method	0 to 300 mm	6.5 µm
208	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Vernier Gauge (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauges, Surface Plate by Comparison Method	0 to 300 mm	15 µm
209	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Thickness Gauge, Thickness Gauge (Jewel Type), Digital Thickness Gauge, Rubber Thickness Gauge (L.C.: 0.001 mm)	Using Gauge Block Set by Comparison Method	0 to 25 mm	1.5 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 33 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
210	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Thickness Gauge, Thickness Gauge (Jewel Type), Digital Thickness Gauge, Rubber Thickness Gauge (L.C.: 0.01 mm)	Using Gauge Block Set by Comparison Method	0 to 100 mm	6.8 µm
211	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Thickness Gauge, Thickness Gauge (Jewel Type), Digital Thickness Gauge, Rubber Thickness Gauge (L.C.: 0.01 mm)	Using Gauge Block Set by Comparison Method	0 to 25 mm	6.8 µm
212	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer, Outside Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set and Long Slip Gauges, Optical Flat by Comparison Method	0 to 150 mm	3 µm
213	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer, Outside Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set, Optical Flat by Comparison Method	0 to 25 mm	1.3 µm
214	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer, Outside Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set and Long Slip Gauges, Optical Flat by Comparison Method	> 150 mm to 300 mm	4.1 µm
215	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer, Outside Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set, Long Slip Gauges, Optical Flat by Comparison Method	> 300 mm to 1000 mm	9.3 µm

This is annexure to 'Certificate of Accreditation' and does not require any signature.



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

34 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
216	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler Gauge	Using Digital Micrometer by Direct Method	0.03 mm to 1 mm	4 µm
217	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler Gauge, Shim	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	> 2 mm to 5 mm	2 µm
218	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler Gauge, Shim	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm) and Comparator Stand by Comparison Method	0.01 mm to 2 mm	1.8 µm
219	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Fillet Welding Gauge (Convex and Concave) - Radius	Using Profile Projector by Comparison Method	0.5 mm to 25 mm	12.4 µm
220	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Foil	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm) and Comparator Stand by Comparison Method	0.01 mm to 12 mm	1.8 µm
221	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Hegman Gauge	Using Dial Indicator (L.C.: 0.1 µm) by Comparison Method	Up to 100 µm	1.4 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 35 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
222	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge - Vernier / Dial / Digital (L.C.: 0.01 mm)	Using Caliper Checker, Surface Plate by Comparison Method	0 to 600 mm	13.6 µm
223	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge - Vernier / Dial / Digital (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauge, Surface Plate by Comparison Method	0 to 600 mm	12 µm
224	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Inside Digital Caliper, Inside Dial Caliper, Inside Pistol Caliper, Inside Groove Caliper (L.C.: 0.01 mm)	Using Gauge Block Set with Slip Gauge Accessories, External Micrometer (L.C.: 1 µm) by Comparison Method	5 mm to 300 mm	8.4 µm
225	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type), Tubular Inside Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set, Long Slip Gauge, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	200 mm to 1000 mm	10.5 µm
226	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type), Tubular Inside Micrometer (L.C.: 0.001 mm)	Using Gauge Block Set, Long Slip Gauge, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	200 mm to 2100 mm	20.6 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 36 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
227	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type), Tubular Inside Micrometer (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauge with Slip Gauge Accessories, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate, External Micrometer (L.C.: 1 µm) by Comparison Method	25 mm to 300 mm	7.5 µm
228	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type), Tubular Inside Micrometer (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauge, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	25 mm to 1000 mm	10.9 µm
229	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type), Tubular Inside Micrometer (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauge, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	25 mm to 2100 mm	21.4 µm
230	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer - Caliper Type (L.C.: 0.001 mm)	Using Gauge Block Set with Slip Gauge Accessories by Comparison Method	5 mm to 100 mm	3 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

37 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
231	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer - Caliper Type (L.C.: 0.01 mm)	Using Gauge Block Set with Slip Gauge Accessories by Comparison Method	5 mm to 300 mm	7.9 µm
232	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever Dial Gauge, Lever Dial Indicator (L.C.: 0.002 mm)	Using Dial Calibration Tester by Comparison Method	0 to 0.8 mm	2.1 µm
233	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever Dial Gauge, Lever Dial Indicator (L.C.: 0.01 mm)	Using Dial Calibration Tester by Comparison Method	0 to 1.6 mm	3.5 µm
234	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever Type Dial Gauge, Lever Type Dial Indicator (L.C.: 0.001 mm)	Using Dial Calibration Tester by Comparison Method	0 to 0.14 mm	1.8 µm
235	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Master Setting Disc - Diameter	Using Gauge Block Set, Long Slip Gauge, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	> 100 mm to 225 mm	2.8 µm
236	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Master Setting Disc - Diameter	Using Gauge Block Set, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	> 25 mm to 100 mm	2 µm

This is annexure to 'Certificate of Accreditation' and does not require any signature.



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

38 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
237	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Master Setting Disc - Diameter	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	3 mm to 25 mm	1.8 µm
238	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Pin Gauge, Measuring Cylindrical Pin	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	0.1 mm to 20 mm	1.8 µm
239	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Scale, Dip Rod Scale, Scale, Ruler (L.C.: 0.5 mm and coarser)	Using Tape and Scale Measuring Machine by Comparison Method	0 to 1000 mm	147.2 µm
240	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Scale, Dip Rod Scale, Scale, Ruler (L.C.: 1 mm and coarser)	Using Tape and Scale Measuring Machine by Comparison Method	0 to 2000 mm	205.5 x Sqrt (L) µm, where L is in meter
241	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Tape, Dip Tape (L.C.: 1 mm)	Using Tape and Scale Measuring Machine by Comparison Method	0 to 100 m	98.6 x Sqrt (L) µm, where L is in meter
242	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Gauge Block Set, Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	> 150 mm to 300 mm	3.5 µm

This is annexure to 'Certificate of Accreditation' and does not require any signature.



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

39 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
243	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Gauge Block Set, Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	> 300 mm to 500 mm	6.4 µm
244	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Gauge Block Set, Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	> 500 mm to 800 mm	8.9 µm
245	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Gauge Block Set, Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	> 800 mm to 1000 mm	10.5 µm
246	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Gauge Block Set, Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	25 mm to 150 mm	2.5 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

40 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
247	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Outside Digital Caliper, Outside Dial Caliper, Outside Pistol Caliper, Outside Groove Caliper (L.C.: 0.01 mm)	Using Gauge Block Set by Comparison Method	0 to 300 mm	8.3 µm
248	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pick Glass Scale (L.C.: 0.1 mm and Coarser) - Scale Length	Using Profile Projector by Comparison Method	0 to 51 mm	31.4 µm
249	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pie Tape, Circumference Tape (L.C.: 0.02 mm and Coarser)	Using Tape and Scale Measuring Machine by Comparison Method	0 to 25.9182 m	90 x Sqrt (L) µm, where L is in meter
250	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pitch Gauge Set - Pitch	Using Profile Projector by Comparison Method	0.05 mm to 7 mm	12.4 µm
251	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pitch Gauge Set, Protractor (Geometric Box) - Angle (All Quadrant's)	Using Profile Projector by Comparison Method	0 ° to 180 °	3 minute of arc



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

41 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
252	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge - Diameter	Using Gauge Block Set, Long Slip Gauge, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	> 100 mm to 225 mm	2.8 µm
253	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge - Diameter	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	1 mm to 25 mm	1.8 µm
254	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge - Diameter	Using Gauge Block Set, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	25 mm to 100 mm	2 µm
255	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Gauge, Plunger Dial Indicator (L.C.: 0.001 mm)	Using Gauge Block Set and Comparator Stand by Comparison Method	0 to 25 mm	2.1 µm
256	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Gauge, Plunger Dial Indicator (L.C.: 0.001 mm)	Using Dial Calibration Tester by Comparison Method	0 to 25 mm	2.5 µm
257	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Gauge, Plunger Dial Indicator (L.C.: 0.01 mm)	Using Dial Calibration Tester by Comparison Method	0 to 25 mm	6.3 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 42 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
258	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Gauge, Plunger Dial Indicator (L.C.: 0.01 mm)	Using Gauge Block Set and Comparator Stand by Comparison Method	0 to 25 mm	7 µm
259	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Indicator, Plunger Dial Gauge (L.C.: 0.0001 mm)	Using Gauge Block Set and Comparator Stand by Comparison Method	0 to 25 mm	1.3 µm
260	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Indicator, Plunger Dial Gauge (L.C.: 0.01 mm)	Using Gauge Block Set and Comparator Stand by Comparison Method	0 to 100 mm	8 µm
261	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial Indicator, Plunger Dial Gauge (L.C.: 0.01 mm)	Using Gauge Block Set and Comparator Stand by Comparison Method	0 to 50 mm	7.5 µm
262	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Radius Gauge Set, Profile Gauge, Template - Radius (Convex & Concave)	Using Profile Projector by Comparison Method	0.5 mm to 25 mm	12.4 µm
263	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Radius Gauge Set, Profile Gauge, Template - Radius (Convex & Concave)	Using Profile Projector by Comparison Method	25 mm to 100 mm	12.4 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

43 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
264	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Sine Bar - Angle	Using Angle Gauges, Dial Indicator (L.C.: 0.1 µm) by Comparison Method	0° to 90°	5.1 second of arc
265	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Sine Bar - Centre Distance	Using Slip Gauge Set, Dial Indicator (L.C.: 0.1 µm) by Comparison Method	Up to 300 mm	5.6 µm
266	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Sine Bar - Parallelism	Using Surface Plate, Dial Indicator (L.C. 0.1µm) by Comparison Method	Up to 300 mm	5.6 µm
267	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge	Using Gauge Block Set, Long Slip Gauge Set by Comparison Method	3 mm to 300 mm	6.9 µm
268	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Step Block, Height Block - Height	Using Gauge Block Set and Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Surface Plate by Comparison Method	0.5 mm to 300 mm	3.5 µm
269	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Surface Plate	Using Precision Level by Comparison Method	Up to 1000 mm X 1000 mm	3 x Sqrt {(L+W) / 125} µm, Where L and W are in mm

This is annexure to 'Certificate of Accreditation' and does not require any signature.



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 44 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
270	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Taper Slot Gauge - Depth (L.C.: 0.01 mm and Coarser)	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	0 to 15 mm	7.2 µm
271	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Taper Slot Gauge - Depth (L.C.: 0.01 mm and Coarser)	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand, Gauge Block Set by Comparison Method	0 to 40 mm	7.2 µm
272	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Test Sieve	Using Profile Projector by Comparison Method	0.025 mm to 4 mm	5 µm
273	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Test Sieve	Using Digital Caliper by Comparison Method	4 mm to 125 mm	20 µm
274	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Ultrasonic Thickness Gauge (L.C.: 0.01 mm and Coarser)	Using Gauge Block Set and Long Slip Gauge Set by Comparison Method	0 to 100 mm	11.2 µm
275	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Ultrasonic Thickness Gauge (L.C.: 0.1 mm)	Using Gauge Block Set and Long Slip Gauge Set by Comparison Method	0 to 300 mm	111.7 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 45 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
276	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Welding Gauge (Leaf or Sheet Type), Set Square (L.C.: 1°)	Using Profile Projector by Comparison Method	0 ° to 180 °	35.23 minute of arc
277	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Welding Gauge - Linear (L.C.: 0.5 mm and Coarser)	Using Profile Projector by Comparison Method	0 to 100 mm	0.144 mm
278	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Width Gauge - Size Variation	Using Gauge Block Set, Long Slip Gauges, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	0.5 mm to 300 mm	3.5 µm
279	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Wire Gauge - Diameter	Using Profile Projector by Comparison Method	0.125 mm to 8.25 mm	12.7 µm
280	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Dial Calibration Tester (L.C.: 0.0002 mm)	Using Gauge Block Set, Comparator Stand by Comparison Method	0 to 25 mm	1.3 µm
281	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Dial Calibration Tester (L.C.: 0.0002 mm)	Using Dial Indicator (L.C.: 0.1 µm) by Comparison Method	0 to 25 mm	1.5 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

46 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
282	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Electronic Height Gauge - Linear (L.C.: 0.0001 mm)	Using Gauge Block Set, Long Slip Gauge, Surface Plate by Comparison Method	0 to 600 mm	7.5 µm
283	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Electronic Height Gauge - Linear (L.C.: 0.0001 mm)	Using Caliper Checker, Surface Plate by Comparison Method	0 to 600 mm	8 µm
284	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Electronic Height Gauge - Squareness (L.C.: 0.0001 mm)	Using Granite Square Master, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Surface Plate by Comparison Method	0 to 600 mm	11.8 µm
285	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	LVDT Probe, Digital Probe, Digital Dial Gauge, Digital Plunger Dial Indicator, Electronic Probe with Display Unit or DRO (L.C.: 0.0001 mm)	Using Gauge Block Set, Comparator Stand by Comparison Method	0 to 25 mm	1.3 µm
286	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Micrometer Head (L.C.: 0.0002 mm)	Using Gauge Block Set, Comparator Stand by Comparison Method	0 to 25 mm	1.3 µm
287	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Micrometer Head (L.C.: 0.0002 mm)	Using Digital Plunger Dial Gauge (L.C.: 0.1 µm), Comparator Stand by Comparison Method	0 to 25 mm	1.5 µm
288	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile Projector - Magnification	Using Gauge Block and Digital Caliper by Comparison Method	10 X to 100 X	1 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

47 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
289	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile Projector, Video Measuring Machine - Angular (L.C.: 0.01 second of arc and Coarser)	Using Standard Angular Gauge Block By Comparison Method	0 ° to 360 °	1.9 minute of arc
290	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile Projector, Video Measuring Machine - Linear (L.C.: 0.001 mm and Coarser)	Using Linear Glass Scale by Comparison Method	0 to 200 mm	4.3 µm
291	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Tape & Scale Calibration Machine, Tape & Scale Calibrator (L.C.: 0.001 mm)	Using Gauge Block Set & Long Slip Gauge Set by Comparison Method	0 to 1000 mm	10.2 µm
292	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Barometer, Atmospheric Pressure Indicator, Atmospheric Pressure Calibrator, Manometer, Barometric Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Pump, Pressure Calibrator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	700 mbar (abs) to 1200 mbar (abs)	0.2 mbar (abs)
293	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder - Absolute Pressure	Using Pneumatic Pressure Source (Nitrogen Gas Cylinder), Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0.014 bar (abs) to 70 bar (abs)	0.05 bar (abs)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 48 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
294	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Manometer - Pneumatic Absolute Pressure	Using Pneumatic Pressure Dead Weight Tester, Digital Multimeter by Direct Method as per DKD-R 6-1	0.014 bar (abs) to 30 bar (abs)	0.003 bar (abs)
295	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Source (Nitrogen Gas Cylinder), Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 40 bar	0.007 bar
296	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 2 bar	0.0003 bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

49 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
297	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 20 bar	0.002 bar
298	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Source (Nitrogen Gas Cylinder), Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 70 bar	0.018 bar
299	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Pressure Module with Indicator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	(-) 40 kPa to 40 kPa	0.005 kPa



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 50 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
300	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Hydraulic Pressure	Using Hydraulic Pressure Dead Weight Tester, Digital Multimeter by Direct Method as per DKD-R 6-1	20 bar to 1200 bar	0.01 % rdg
301	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Hydraulic Pressure	Using Hydraulic Pressure Dead Weight Tester, Digital Multimeter by Direct Method as per DKD-R 6-1	3 bar to 35 bar	0.01 % rdg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

51 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
302	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Pneumatic Low Pressure	Using Low Pressure Pump, Low Pressure Calibrator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 400 mbar	0.05 mbar
303	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Dead Weight Tester, Digital Multimeter by Direct Method as per DKD-R 6-1	3 bar to 35 bar	0.01 %rdg
304	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Hydraulic Pressure	Using Hydraulic Pressure Pump, Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 350 bar	0.047 bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 52 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
305	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Hydraulic Pressure	Using Hydraulic Pressure Pump, Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 700 bar	0.085 bar
306	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Vacuum Gauge, Vacuum Indicator, Vacuum Calibrator, Vacuum Controller, Vacuum Transmitter, Vacuum Transducer, Manometer, Vacuum Recorder, Vacuum Switch with Electrical Output - Pneumatic Low Vacuum	Using Low Pressure Pump, Low Pressure Calibrator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	(-) 400 mbar to 0 mbar	0.05 mbar
307	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Vacuum Gauge, Vacuum Indicator, Vacuum Calibrator, Vacuum Controller, Vacuum Transmitter, Vacuum Transducer, Vacuum Recorder, Vacuum Switch with Electrical Output - Pneumatic Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	(-) 1 bar to 0 bar	0.0006 bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 53 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
308	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Vacuum Gauge, Vacuum Indicator, Vacuum Calibrator, Vacuum Controller, Vacuum Transmitter, Vacuum Transducer, Vacuum Recorder, Vacuum Switch with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure / Vacuum Dead Weight Tester, Digital Multimeter by Direct Method as per DKD-R 6-1	(-) 100 mbar to (-) 1000 mbar	0.01 %rdg
309	MECHANICAL-PRESSURE INDICATING DEVICES	Analog, Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Manometer - Pneumatic Absolute Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0.014 bar (abs) to 21 bar (abs)	0.0032 bar (abs)
310	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog - Differential Pressure Gauge, Pressure Gauge, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Low Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 490 Pa	0.38 Pa



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 54 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
311	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog - Differential Pressure Gauge, Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Low Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 1034 Pa	0.35 Pa
312	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog - Differential Pressure, Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Low Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 2068 Pa	1.42 Pa
313	MECHANICAL-TORQUE GENERATING DEVICES	Indicating Torque Tool (Type I - Class A to Class C), Setting Torque Tool (Type II - Class A to Class C, Class G)	Using Torque Wrench Calibration System with Torque Sensors and Digital Torque Indicator as per ISO 6789-1:2017 and ISO 6789-2:2017	100 Nm to 400 Nm	0.6 % rdg
314	MECHANICAL-TORQUE GENERATING DEVICES	Indicating Torque Tool (Type I - Class A to Class C), Setting Torque Tool (Type II - Class A to Class C, Class G)	Using Torque Wrench Calibration System with Torque Sensors and Digital Torque Indicator as per ISO 6789-1:2017 and ISO 6789-2:2017	20 Nm to 100 Nm	0.9 % rdg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 55 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
315	MECHANICAL-TORQUE GENERATING DEVICES	Indicating Torque Tool (Type I - Class A to Class C), Setting Torque Tool (Type II - Class A to Class C, Class G)	Using Torque Wrench Calibration System with Torque Sensors and Digital Torque Indicator as per ISO 6789-1:2017 and ISO 6789-2:2017	400 Nm to 2000 Nm	0.61 % rdg
316	MECHANICAL-TORQUE GENERATING DEVICES	Indicating Torque Tool (Type I - Class A to Class E), Setting Torque Tool (Type II - Class A to Class G)	Using Torque Wrench Calibration System with Torque Sensors and Digital Torque Indicator as per ISO 6789-1:2017 and ISO 6789-2:2017	0.5 Nm to 5 Nm	3.08 % rdg
317	MECHANICAL-TORQUE GENERATING DEVICES	Indicating Torque Tool (Type I - Class A to Class E), Setting Torque Tools (Type II - Class A to Class G)	Using Torque Wrench Calibration System with Torque Sensors and Digital Torque Indicator as per ISO 6789-1:2017 and ISO 6789-2:2017	5 Nm to 20 Nm	1.22 % rdg
318	MECHANICAL-VOLUME	Dispenser	Using Weighing Balance (Readability : 0.1 mg), Distilled Water by Gravimetric Method as per ISO 8655-6:2022(E) & ISO 8655-5:2022	> 50 ml to 100 ml	0.08 ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 56 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
319	MECHANICAL-VOLUME	Measuring Jug, Beaker	Using Weighing Balance (Readability : 2 mg), Distilled Water by Gravimetric Method as per ISO 4787:2021(E)	> 5000 ml to 10000 ml	3 ml
320	MECHANICAL-VOLUME	Measuring Jug, Measuring Cylinder, Beaker	Using Weighing Balance (Readability : 10 mg), Distilled Water by Gravimetric Method as per ISO 4787:2021(E)	> 1000 ml to 2000 ml	0.5 ml
321	MECHANICAL-VOLUME	Measuring Jug, Measuring Cylinder, Beaker	Using Weighing Balance (Readability : 10 mg), Distilled Water by Gravimetric Method as per ISO 4787:2021(E)	> 2000 ml to 5000 ml	1.5 ml
322	MECHANICAL-VOLUME	Measuring Jug, Measuring Cylinder, Burette, Flask, Beaker	Using Weighing Balance (Readability : 0.01 mg), Distilled Water by Gravimetric Method as per ISO 4787:2021(E)	> 50 ml to 100 ml	0.3 ml
323	MECHANICAL-VOLUME	Measuring Jug, Measuring Cylinder, Burette, Flask, Beaker	Using Weighing Balance (Readability : 0.01 mg), Distilled Water by Gravimetric Method as per ISO 4787:2021(E)	1 ml to 50 ml	3 µl



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 57 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
324	MECHANICAL-VOLUME	Measuring Jug, Measuring Cylinder, Flask, Beaker	Using Weighing Balance (Readability : 0.1 mg), Distilled Water by Gravimetric Method as per ISO 4787:2021(E)	> 100 ml to 1000 ml	0.3 ml
325	MECHANICAL-VOLUME	Non Medical Syringe (Delivery of Liquid), Micro Pipette	Using Weighing Balance (Readability : 0.01 mg), Distilled Water by Gravimetric Method as per ISO 8655-6:2022(E)	> 1000 µl to 10000 µl	50 µl
326	MECHANICAL-VOLUME	Non Medical Syringe (Delivery of Liquid), Micro Pipette	Using Weighing Balance (Readability : 0.01 mg), Distilled Water by Gravimetric Method as per ISO 8655-6:2022(E)	> 100 µl to 1000 µl	2 µl
327	MECHANICAL-VOLUME	Non Medical Syringe (Delivery of Liquid), Micro Pipette	Using Weighing Balance (Readability : 0.001 mg), Distilled Water by Gravimetric Method as per ISO 8655-6:2022(E)	10 µl to 100 µl	0.4 µl
328	MECHANICAL-VOLUME	Piston Operated Burette, Dispenser	Using Weighing Balance (Readability : 0.1 mg), Distilled Water by Gravimetric Method as per ISO 8655-6:2022(E), ISO 8655-3:2022 & ISO 8655-5:2022	> 10 ml to 50 ml	0.05 ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 58 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
329	MECHANICAL-VOLUME	Piston Operated Burette, Dispenser	Using Weighing Balance (Readability : 0.01 mg), Distilled Water by Gravimetric Method as per ISO 8655-6:2022(E), ISO 8655-3:2022 & ISO 8655-5:2022	1 ml to 10 ml	0.03 ml
330	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.0001 g)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 5 kg	0.9 mg
331	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.0001 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 6 g	0.005 mg
332	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.001 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 50 g	0.005 mg
333	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.01 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 200 g	0.027 mg
334	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 1 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 1 kg	1.1 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

59 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
335	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 2 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 50 kg	20 mg
336	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class II & Coarser (Readability : 10 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 6 kg	11 mg
337	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class III & Coarser (Readability : 5g)	Using F1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 75 kg	4 g
338	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class IIII (Readability : 20 g)	Using F1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 200 kg	20 g
339	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class IIII (Readability : 20 g)	Using F1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 500 kg	20 g
340	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	1 mg	0.0006 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

60 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
341	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	1 g	0.0006 mg
342	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	1 mg	0.0005 mg
343	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	10 g	0.003 mg
344	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	10 mg	0.0005 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

61 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
345	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	10 mg	0.0006 mg
346	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Weighing Balance (Readability : 0.01 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	100 g	0.014 mg
347	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	100 mg	0.0005 mg
348	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	100 mg	0.0006 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

62 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
349	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	2 g	0.0006 mg
350	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	2 mg	0.0005 mg
351	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	2 mg	0.0006 mg
352	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	20 g	0.003 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

63 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
353	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	20 mg	0.0005 mg
354	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	20 mg	0.0006 mg
355	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Weighing Balance (Readability : 0.01 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	200 g	0.021 mg
356	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	200 mg	0.0005 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

64 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
357	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	200 mg	0.0006 mg
358	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	5 g	0.0015 mg
359	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	5 mg	0.0005 mg
360	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	5 mg	0.0006 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

65 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
361	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	50 g	0.005 mg
362	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	50 mg	0.0005 mg
363	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	50 mg	0.0006 mg
364	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Subdivision Method (ABBA Cycle) as per OIML R 111-1:2004	500 mg	0.0005 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

66 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
365	MECHANICAL-WEIGHTS	Accuracy Class E1 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 mg) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	500 mg	0.0006 mg
366	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	1 kg	0.3 mg
367	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.002 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	10 kg	3 mg
368	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	2 kg	0.4 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

67 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
369	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.002 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	20 kg	7 mg
370	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	5 kg	1 mg
371	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.002 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	50 kg	17 mg
372	MECHANICAL-WEIGHTS	Accuracy Class E2 & Coarser	Using E1 Class Weight and Mass Comparator (Readability : 0.0001 g) by Substitution Method (ABBA Cycle) as per OIML R 111-1:2004	500 g	0.2 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

68 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
373	THERMAL-SPECIFIC HEAT & HUMIDITY	Portable Temperature Indicator / Recorder / Logger	Using Standard RTD, Precision Temperature Scanner with Data Monitoring System and Temperature Chamber by Comparison Method	(-) 30 °C to 85 °C	0.13 °C
374	THERMAL-SPECIFIC HEAT & HUMIDITY	Temperature & Humidity Test / Calibration Chamber, Environment Test Chamber / Room, Climatic Test / Calibration Chamber / Room @ 23°C to 50°C - Multi Position (Minimum 9 Sensors)	Using Temperature & Humidity Recorder with Data Monitoring System by Comparison Method	10 % RH to 95 % RH	1.86 % RH
375	THERMAL-SPECIFIC HEAT & HUMIDITY	Temperature & Humidity Test / Calibration Chamber, Environment Test Chamber / Room, Climatic Test / Calibration Chamber / Room @ 25°C - Multi Position (Minimum 9 Sensors)	Using Temperature & Humidity Recorder with Data Monitoring System by Comparison Method	10 % RH to 95 % RH	1.7 % RH



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 69 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
376	THERMAL-SPECIFIC HEAT & HUMIDITY	Temperature & Humidity Test / Calibration Chamber, Environment Test Chamber / Room, Climatic Test / Calibration Chamber / Room @ 50%RH - Multi Position (Minimum 9 Sensors)	Using Temperature & Humidity Recorder with Data Monitoring System by Comparison Method	5 °C to 50 °C	0.85 °C
377	THERMAL-SPECIFIC HEAT & HUMIDITY	Thermo hygrometer with Internal Sensor, Digital Humidity Meter, Humidity Sensor with Indicator, Humidity Transmitter with Indicator, Temperature & Humidity Recorder @ 50% RH	Using Standard Digital RH & Temperature Indicator with Sensor, Precision Digital Multimeter, Stable Relative Humidity & Temperature Chamber by Comparison Method	5 °C to 50 °C	0.2 °C
378	THERMAL-SPECIFIC HEAT & HUMIDITY	Thermo hygrometer with Internal Sensor, Digital Humidity Meter, Temperature Humidity Sensor with Indicator, Temperature Humidity Transmitter, Temperature & Humidity Recorder @ 23°C to 50°C	Using Digital RH & Temperature Indicator with Sensor, Precision Digital Multimeter and Stable Relative Humidity & Temperature Chamber by Comparison Method	10 % RH to 95 % RH	1.88 % RH



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 70 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
379	THERMAL-SPECIFIC HEAT & HUMIDITY	Thermo hygrometer with Internal Sensor, Digital Humidity Meter, Temperature Humidity Sensor with Indicator, Temperature Humidity Transmitter, Temperature & Humidity Recorder @ 25 °C	Using Digital RH & Temperature Indicator with Sensor, Precision Digital Multimeter, Stable Relative Humidity & Temperature Chamber by Comparison Method	10 % RH to 95 % RH	0.7 % RH
380	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Dry Block Calibrator, Temperature Furnace -Single Position	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner by Comparison Method	280 °C to 660 °C	0.12 °C
381	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Dry Block Calibrator, Temperature Furnace -Single Position	Using S Type Thermocouple Sensor, Precision Temperature Scanner by Comparison Method	660 °C to 1200 °C	1.72 °C
382	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Liquid Bath Calibrator, Temperature Dry Block Calibrator - Single Position	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner by Comparison Method	140 °C to 280 °C	0.02 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 71 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
383	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Liquid Bath Calibrator, Temperature Water Bath, Temperature Dry Block Calibrator - Single Position	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner by Comparison Method	(-) 95 °C to 140 °C	0.01 °C
384	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using Non Contact Infrared Thermometer, Precision Digital Multimeter, Black Body Source by Comparison Method	> 100 °C to 500 °C	2.32 °C
385	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using Standard RTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Black Body Source by Comparison Method	(-) 20 °C to 100 °C	0.58 °C
386	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using Non Contact Infrared Thermometer, Precision Digital Multimeter, Black Body Source by Comparison Method	(-) 20 °C to 100 °C	1 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 72 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
387	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using S - Type Thermocouple Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Black Body Source by Comparison Method	> 100 °C to 500 °C	2.12 °C
388	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.99)	Using S Type Thermocouple Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Black Body Source by Comparison Method	> 500 °C to 1200 °C	5.66 °C
389	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.99)	Using Non Contact Infrared Thermometer, Precision Digital Multimeter, Black Body Source by Comparison Method	> 500 °C to 1200 °C	7.36 °C
390	THERMAL-TEMPERATURE	Liquid-In-Glass Thermometer	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Liquid Bath by Comparison Method	(-) 80 °C to 0 °C	0.62 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 73 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
391	THERMAL-TEMPERATURE	Liquid-In-Glass Thermometer	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Liquid Bath by Comparison Method	0 °C to 150 °C	0.62 °C
392	THERMAL-TEMPERATURE	Liquid-In-Glass Thermometer	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Liquid Bath by Comparison Method	150 °C to 280 °C	0.62 °C
393	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Liquid Nitrogen bath by Comparison Method	(-) 196 °C	0.4 °C
394	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Dry Block Calibrator by Comparison Method	280 °C to 660 °C	0.16 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 74 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
395	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Dry Block Calibrator by Comparison Method	(-) 95 °C to 140 °C	0.03 °C
396	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Liquid Temperature Calibrator by Comparison Method	140 °C to 280 °C	0.02 °C
397	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using S Type Thermocouple Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Temperature Furnace Calibrator by Comparison Method	660 °C to 1200 °C	1.69 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

75 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.01 Lead / Lag to UPF, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Calibrator by Direct / Comparison Method	0.04 Wh to 32 kWh	0.0047 % to 0.4 %
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.01 Lead / Lag to UPF, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 Wh to 38.4 kWh	0.0047 % to 1.6 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 W to 32 kW	0.0047 % to 0.4 %
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 W to 38.4 kW	0.0047 % to 1.6 %
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VAh to 32 kVAh	0.0047 % to 0.03 %
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	4 kVAh to 38.4 kVAh	0.0047 % to 0.0162 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 76 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VA to 32 kVA	0.0047 % to 0.035 %
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 VA to 38.4 kVA	0.0047 % to 0.0162 %
9	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VARh to 32 kVARh	0.03 % to 0.4 %
10	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 VARh to 38.4 kVARh	0.0047 % to 1.6 %
11	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.04 VAR to 32 kVAR	0.005 % to 0.4 %
12	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	1Ø, 2 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	40 VAR to 38.4 kVAR	0.0047 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

77 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
13	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.1 Lead / Lag to UPF, 40 V to 320 V, 10 mA to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 Wh to 115.2 kWh	0.008 % to 0.16 %
14	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 10 mA, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 W to 9.6 W	0.008 % to 0.8 %
15	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 10 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.012 W to 115.2 kW	0.008 % to 1.6 %
16	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 10 mA to 120 A, 0.1 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VARh to 115.2 kVARh	0.008 % to 0.16 %
17	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 10 mA, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.19 VAR to 9.6 VAR	0.008 % to 0.8 %
18	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 3 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 10 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.012 VAR to 115.2 kVAR	0.008 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

78 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
19	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Active Energy @ (45 Hz to 65 Hz, 0.01 Lead / Lag to UPF, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 Wh to 96 kWh	0.0047 % to 0.4 %
20	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 W to 96 kW	0.031 % to 0.4 %
21	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VA to 96 kVA	0.0047 % to 0.031 %
22	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VARh to 96 kVARh	0.0047 % to 0.4 %
23	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, AC Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VAR to 96 kVAR	0.0047 % to 0.4 %
24	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Active Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 Wh to 115.2 kWh	0.0047 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

79 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
25	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Active Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 W to 115.2 kW	0.0047 % to 1.6 %
26	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 1 mA to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.12 VAh to 96 kVAh	0.0047 % to 0.031 %
27	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Apparent Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	12 kVAh to 115.2 kVAh	0.0047 % to 0.0162 %
28	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Apparent Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 VA to 115.2 kVA	0.0047 % to 0.0162 %
29	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Reactive Energy @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 VARh to 115.2 kVARh	0.0047 % to 1.6 %
30	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	3Ø, 4 Wire, Reactive Power @ (45 Hz to 65 Hz, 40 V to 320 V, 100 A to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	120 VAR to 115.2 kVAR	0.0047 % to 1.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 80 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
31	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 10 Hz to 1 kHz	Using 8½ Digit Multimeter with MPC by Comparison Method	30 µA to 100 mA	0.11 % to 0.053 %
32	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 10 Hz to 1 kHz	Using 8½ Digit Multimeter by Direct Method	30 µA to 100 mA	0.11 % to 0.053 %
33	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 45 Hz to 65 Hz	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	0.001 A to 100 A	0.045 % to 0.004 %
34	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 45 Hz to 65 Hz	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	100 A to 120 A	0.004 % to 0.01 %
35	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 1 kHz	Using 8½ Digit Multimeter by Direct Method	100 mA to 1 A	0.053 % to 0.094 %
36	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 1 kHz	Using 8½ Digit Multimeter with MPC by Comparison Method	100 mA to 1 A	0.053 % to 0.094 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

81 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
37	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 5 kHz	Using 8½ Digit Multimeter by Direct Method	1 A to 20 A	0.094 % to 0.12 %
38	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 5 kHz	Using 8½ Digit Multimeter with MPC by Comparison Method	1 A to 20 A	0.094 % to 0.12 %
39	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC High Current @ 50 Hz	Using High Current to Low Current Conversion, STD Current Transformer & Low Current Measured at STD Meter with High Current Source by Direct / Comparison Method	100 A to 6000 A	0.065 % to 0.26 %
40	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC High Voltage @ 50 Hz	Using High Voltage Divider with kV Meter with HV Source by Direct / Comparison Method	1 kV to 100 kV	1.46 % to 1.47 %
41	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter by Direct Method	1 mV to 100 mV	0.5 % to 0.018 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

82 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
42	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter with MPC by Comparison Method	1 mV to 100 mV	0.54 % to 0.019 %
43	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter by Direct Method	100 mV to 1000 V	0.018 % to 0.02 %
44	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 20 Hz to 10 kHz	Using 8½ Digit Multi Meter with MPC by Comparison Method	100 mV to 1000 V	0.019 % to 0.02 %
45	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 45 Hz to 65 Hz	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	30 V to 480 V	0.004 % to 0.002 %
46	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance @ 1 kHz	Using Precision LCR Meter with Decade Capacitance Box by Direct Method / Substitution Method	100 pF to 100 µF	0.1 % to 0.14 %
47	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance @ 100 Hz	Using Precision LCR Meter with MPC by Direct Method / Substitution Method	1 mF to 100 mF	0.14 % to 0.19 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

83 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
48	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Harmonics @ (50 Hz, 40 V to 240 V, 0.05 A to 10 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	1st order to 39th order	0.5 %
49	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Inductance @ 1 kHz	Using Precision LCR Meter with Decade Inductance Box by Direct Method / Substitution Method	100 µH to 10 H	0.1 % to 0.11 %
50	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Power Factor @ (50 Hz, 40 V to 480 V, 0.01 A to 100 A)	Using 3Ø Energy Reference Standard with Power Calibrator by Direct / Comparison Method	(-) 1 PF to 1 PF	0.00001 PF
51	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Resistance @ 1 kHz	Using Precision LCR Meter with Decade Resistance Box by Direct Method / Substitution Method	1 ohm to 10 kohm	0.1 % to 0.08 %
52	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Active Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mWh to 36 kWh	0.071 % to 1.61 %
53	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Active Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mW to 36 kW	0.071 % to 1.61 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

84 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
54	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Reactive Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mVArh to 36 kVArh	0.071 % to 1.61 %
55	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	1Ø, 2 Wire, AC Reactive Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	40 mVAr to 36 kVAr	0.071 % to 1.61 %
56	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 3 Wire, AC Active Energy @ (50 Hz, 63.5 V, 1 mA to 10 A, 0.5 Lag / 0.8 Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	95.25 mWh to 1.905 kWh	0.071 % to 0.061 %
57	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 3 Wire, AC Reactive Energy @ (50 Hz, 63.5 V, 1 mA to 10 A, 0.5 Lag / 0.8 Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	95.25 mVArh to 1.905 kVArh	0.071 % to 0.061 %
58	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Active Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mWh to 108 kWh	0.071 % to 1.61 %
59	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Active Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mW to 108 kW	0.071 % to 1.61 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

85 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
60	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Reactive Energy @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mVArh to 108 kVArh	0.071 % to 1.61 %
61	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	3Ø, 4 Wire, AC Reactive Power @ (45 Hz to 65 Hz, 40 V to 300 V, 1 mA to 120 A, 0.01 Lag / Lead to UPF)	Using 3Ø Power Calibrator by Direct Method	120 mVAr to 108 kVAr	0.071 % to 1.61 %
62	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 45 Hz to 65 Hz	Using 3Ø Power Calibrator by Direct Method	1 mA to 120 A	0.034 % to 0.025 %
63	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz	Using Multi Product Calibrator with 50 Turns Current Coil by Direct Method	20 A to 1000 A	0.5 % to 0.55 %
64	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multi Product Calibrator by Direct Method	1 mA to 100 mA	0.14 % to 0.087 %
65	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multi Product Calibrator by Direct Method	100 mA to 20 A	0.087 % to 0.21 %
66	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multi Product Calibrator by Direct Method	30 µA to 1 mA	0.58 % to 0.14 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 86 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
67	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Inductance @ 1 kHz	Using Decade Inductance Box by Direct Method	100 μ H to 1000 mH	0.6 % to 0.18 %
68	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	1 mV to 3 mV	1.7 % to 0.4 %
69	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	1 V to 1000 V	0.04 % to 0.063 %
70	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	100 mV to 1 V	0.06 % to 0.04 %
71	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50 Hz to 10 kHz	Using Multi Product Calibrator by Direct Method	3 mV to 100 mV	0.4 % to 0.06 %
72	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	Using Decade Capacitance Box by Direct Method	1 μ F to 100 μ F	1.8 % to 0.5 %
73	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	Using Decade Capacitance Box by Direct Method	100 pF to 1 μ F	0.7 % to 1.8 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 87 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
74	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	1 mohm to 100 mohm	0.056 % to 0.026 %
75	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	10 µohm to 50 µohm	0.087 % to 0.057 %
76	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	100 mohm to 1 ohm	0.011 % to 0.021 %
77	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter and MPC by V/I Method	50 µohm to 1 mohm	0.057 % to 0.056 %
78	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter by Direct Method	1 µA to 100 mA	0.05 % to 0.005 %
79	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter with MPC by Comparison Method	1 µA to 100 mA	0.053 % to 0.005 %
80	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter by Direct Method	1 A to 20 A	0.022 % to 0.047 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 88 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
81	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter with MPC by Comparison Method	1 A to 20 A	0.022 % to 0.057 %
82	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter by Direct Method	100 mA to 1 A	0.005 % to 0.022 %
83	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8½ Digit Multimeter with MPC by Comparison Method	100 mA to 1 A	0.005 % to 0.022 %
84	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC High Voltage	Using High Voltage Divider with KV Meter with HV Source by Comparison Method	1 kV to 100 kV	2.21 % to 1.53 %
85	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter by Direct Method	0.1 mV to 1 mV	0.12 % to 0.014 %
86	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter with MPC by Comparison Method	0.1 mV to 1 mV	0.13 % to 0.016 %
87	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter by Direct Method	1 mV to 10 V	0.014 % to 0.0005 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 89 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
88	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter with MPC by Comparison Method	1 mV to 10 V	0.016 % to 0.0005 %
89	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter by Direct Method	10 V to 1000 V	0.0005 % to 0.0007 %
90	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8½ Digit Multimeter with MPC by Comparison Method	10 V to 1000 V	0.0005 % to 0.0007 %
91	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	1 ohm to 100 kohm	0.0027 % to 0.001 %
92	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	1 ohm to 100 kohm	0.0027 % to 0.001 %
93	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	10 Mohm to 100 Mohm	0.004 % to 0.017 %
94	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	10 Mohm to 100 Mohm	0.0025 % to 0.017 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 90 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
95	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	100 kohm to 10 Mohm	0.001 % to 0.004 %
96	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	100 Mohm to 1000 Mohm	0.017 % to 0.1 %
97	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter by Direct Method	100 kohm to 10 Mohm	0.001 % to 0.0025 %
98	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 2 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	100 Mohm to 1000 Mohm	0.017 % to 0.1 %
99	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter by Direct Method	1 mohm to 100 mohm	0.055 % to 0.011 %
100	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with Resistance Box by Substitution Method	1 mohm to 100 mohm	5.8 % to 0.034 %
101	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	1 ohm to 100 kohm	0.021 % to 0.001 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 91 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
102	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter by Direct Method	1 ohm to 100 kohm	0.021 % to 0.001 %
103	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with Resistance Box by Substitution Method	10 µohm to 1 mohm	5 % to 5.8 %
104	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter by Direct Method	100 mohm to 1 ohm	0.011 % to 0.021 %
105	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance - 4 Wire	Using 8½ Digit Multimeter with MPC by Substitution Method	100 mohm to 1 ohm	0.034 % to 0.021 %
106	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator by Direct Method	1 µA to 10 µA	2.34 % to 0.35 %
107	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator by Direct Method	10 µA to 100 mA	0.35 % to 0.016 %
108	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator by Direct Method	100 mA to 20 A	0.016 % to 0.14 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 92 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
109	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multi Product Calibrator with 50 Turns Current Coil by Direct Method	20 A to 1000 A	0.4 %
110	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Decade Megaohmbox Standard by Direct Method	100 kohm to 1000 Mohm	0.1 % to 0.21 %
111	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance @ 1 kV	Using High Resistance Standard Box by Direct Method	10 Tohm	5.4 %
112	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance @ Up to 5 kV	Using Decade Mega Ohm Box Standard by Direct Method	10 Gohm to 1 Tohm	2.6 % to 4.3 %
113	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multi Product Calibrator by Direct Method	1 mV to 100 mV	0.355 % to 0.01 %
114	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multi Product Calibrator by Direct Method	100 mV to 1000 V	0.01 % to 0.008 %
115	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 2 Wire	Using Multi Product Calibrator by Direct Method	1 ohm to 10 kohm	0.07 % to 0.01 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 93 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
116	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 2 Wire	Using Multi Product Calibrator by Direct Method	10 Mohm to 1000 Mohm	0.07 % to 1.74 %
117	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 2 Wire	Using Multi Product Calibrator by Direct Method	10 kohm to 10 Mohm	0.01 % to 0.07 %
118	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 4 Wire	Using Multi Product Calibrator by Direct Method	1 mohm to 1 ohm	6 % to 0.072 %
119	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 4 Wire	Using Multi Product Calibrator by Direct Method	1 ohm to 10 kohm	0.072 % to 0.01 %
120	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance - 4 Wire	Using Low Resistance Standard by Direct Method	10 µohm to 1 ohm	0.64 % to 0.063 %
121	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge CT Mode - Phase Error (Min) Secondary Input - 1 A & 5 A	Using CT - PT Bridge by Comparison Method	120 % to 1 %	0.7 minute to 1.2 minute
122	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge CT Mode - Ratio Error (%) Secondary Input - 1 A & 5 A	Using CT - PT Bridge by Comparison Method	120 % to 1 %	0.013 % to 0.024 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 94 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
123	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge PT Mode - Phase Error (Min) Secondary Input - 63.5 V & 110 V	Using CT - PT Bridge by Comparison Method	120 % to 80 %	0.8 minute to 0.9 minute
124	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT - VT Comparator / Bridge PT Mode - Ratio Error (%) Secondary Input - 63.5 V & 110 V	Using CT - PT Bridge by Comparison Method	120 % to 80 %	0.015 % to 0.016 %
125	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Phase Error (1 A to 3200 A with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	1.18 minute to 2.3 minute
126	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Phase Error (3200 A to 6000 A Primary with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	1.18 minute to 2.3 minute
127	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Ratio Error (1 A to 3200 A with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	0.025 % to 0.04 %
128	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer - Ratio Error (3200 A to 6000 A Primary with Secondary - 1 A & 5 A)	Using Standard CT & Comparator with Source Primary Injection by Comparison Method	120 % to 1 %	0.025 % to 0.04 %
129	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Current Transformer Burden Box	Using Power Network Analyzer with Source by Direct Method	1 VA to 30 VA	0.07 % to 0.06 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 95 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
130	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Potential Transformer / Capacitor Voltage Transformer - Phase Error (220 V / 110 V to 33 kV / 110 V with Secondary - 63.5 V & 110 V)	Using Standard PT & Comparator with Source Primary Injection by Comparison Method	120 % to 80 %	2.18 % to 2.3 %
131	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Potential Transformer / Capacitor Voltage Transformer - Ratio Error (220 V / 110 V to 33 kV / 110 V with Secondary - 63.5 V & 110 V)	Using Standard PT & Comparator with Source Primary Injection by Comparison Method	120 % to 80 %	0.07 %
132	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Potential Transformer Burden Box	Using Power Network Analyzer with Source by Direct Method	1.25 VA to 200 VA	0.07 % to 0.06 %
133	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	Turn Ratio	Using 6½ Digit DMM & 8½ Digit Multimeter with Turn Ratio Standard by V/V Method	0.8 Turn to 2000 Turn	0.09 % to 0.23 %
134	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Conductivity Meter (1 µS to 1000 mS)	Using Decade Resistance Box by Simulation Method	1 ohm to 1 Mohm	0.68 % to 0.6 %
135	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	pH Meter (0 pH to 14 pH)	Using Multi Product Calibrator by Simulation Method	(-) 420 mV to 420 mV	0.3 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 96 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
136	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT 100)	Using 8½ Digit Multimeter with MPC by Substitution Method	(-) 200 °C to 850 °C	0.02 °C
137	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT 100, PT 1000)	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 850 °C	0.018 °C
138	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT 1000)	Using 8½ Digit Multimeter with MPC by Substitution Method	(-) 200 °C to 630 °C	0.02 °C
139	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple J - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 210 °C to 1200 °C	0.05 °C
140	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple B - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	600 °C to 1820 °C	0.124 °C
141	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple B - Type	Using 8½ Digit Multimeter by Direct Method	600 °C to 1820 °C	0.12 °C
142	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple E - Type	Using 8½ Digit Multimeter by Direct Method	(-) 250 °C to 1000 °C	0.08 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 97 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
143	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple E - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 250 °C to 1000 °C	0.09 °C
144	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple J - Type	Using 8½ Digit Multimeter by Direct Method	(-) 210 °C to 1200 °C	0.046 °C
145	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple K - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 1372 °C	0.051 °C
146	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple K - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 1372 °C	0.055 °C
147	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple L - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 900 °C	0.037 °C
148	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple L - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 900 °C	0.039 °C
149	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple N - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 1300 °C	0.073 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 98 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
150	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple N - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 1300 °C	0.079 °C
151	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple R - Type	Using 8½ Digit Multimeter by Direct Method	50 °C to 1767 °C	0.066 °C
152	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple R - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	50 °C to 1767 °C	0.08 °C
153	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple S - Type	Using 8½ Digit Multimeter by Direct Method	50 °C to 1767 °C	0.075 °C
154	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple S - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	50 °C to 1767 °C	0.09 °C
155	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple T - Type	Using 8½ Digit Multimeter by Direct Method	(-) 250 °C to 400 °C	0.114 °C
156	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple T - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 250 °C to 400 °C	0.124 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

99 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
157	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple U - Type	Using 8½ Digit Multimeter by Direct Method	(-) 200 °C to 600 °C	0.045 °C
158	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Thermocouple U - Type	Using 8½ Digit Multimeter with MPC by Comparison Method	(-) 200 °C to 600 °C	0.05 °C
159	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD (PT 100)	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 800 °C	0.03 °C
160	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD (PT 1000)	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 630 °C	0.03 °C
161	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple B - Type	Using Multi Product Calibrator by Direct Method	600 °C to 1820 °C	0.22 °C
162	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple E - Type	Using Multi Product Calibrator by Direct Method	(-) 250 °C to 1000 °C	0.23 °C
163	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple J - Type	Using Multi Product Calibrator by Direct Method	(-) 210 °C to 1200 °C	0.15 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 100 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
164	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple K - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 1372 °C	0.141 °C
165	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple L - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 900 °C	0.243 °C
166	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple N - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 1300 °C	0.19 °C
167	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple R - Type	Using Multi Product Calibrator by Direct Method	50 °C to 1767 °C	0.265 °C
168	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple S - Type	Using Multi Product Calibrator by Direct Method	50 °C to 1767 °C	0.265 °C
169	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple T - Type	Using Multi Product Calibrator by Direct Method	(-) 250 °C to 400 °C	0.343 °C
170	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Thermocouple U - Type	Using Multi Product Calibrator by Direct Method	(-) 200 °C to 600 °C	0.29 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 101 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
171	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter by Direct Method	1 Hz to 50 Hz	0.0002 % to 0.00013 %
172	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter with MPC by Comparison Method	1 Hz to 50 Hz	0.0007 % to 0.00015 %
173	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter by Direct Method	100 kHz to 2 MHz	0.0001 % to 0.00015 %
174	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter with MPC by Comparison Method	100 kHz to 2 MHz	0.00011 % to 0.00013 %
175	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter by Direct Method	50 Hz to 100 kHz	0.00013 % to 0.0001 %
176	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Precision Frequency Universal Counter with MPC by Comparison Method	50 Hz to 100 kHz	0.00015 % to 0.00011 %
177	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time	Using Time Interval Meter by Comparison Method	1 s to 9900 s	0.12 s to 0.53 s



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 102 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
178	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time	Using Time Interval Meter by Comparison Method	36000 s to 86400 s	2.07 s to 7.08 s
179	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time	Using Time Interval Meter by Comparison Method	9900 s to 36000 s	0.53 s to 2.07 s
180	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using Multi Product Calibrator by Direct Method	1 Hz to 50 Hz	0.59 % to 0.012 %
181	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using Multi Product Calibrator by Direct Method	50 Hz to 2 MHz	0.012 % to 0.03 %
182	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 100 RPM to 5000 RPM	3.1 RPM
183	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 5000 RPM to 60000 RPM	16.02 RPM
184	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 60000 RPM to 99950 RPM	20.2 RPM
185	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Indicator with Sensor (Non Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	6 RPM to 100 RPM	1 RPM



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 103 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
186	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Sensor with Indicator, Rotation Meter with Indicator (Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	> 1000 RPM to 11000 RPM	4.01 RPM
187	MECHANICAL-ACCELERATION AND SPEED	Tachometer, RPM Sensor with Indicator, Rotation Meter with Indicator (Contact Type)	Using Tachometer Calibrator, Precision Tachometer by Comparison Method	6 RPM to 1000 RPM	1.01 RPM
188	MECHANICAL-ACOUSTICS	Sound Level Meter, dB Meter, Acoustic Meter, Noise Meter @ 1 kHz	Using Sound Level Calibrator by Comparison Method	94 dB & 114 dB	0.51 dB
189	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge - Vernier / Dial / Digital (L.C.: 0.01 mm)	Using Caliper Checker, Surface Plate by Comparison Method	0 to 600 mm	13.6 µm
190	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge - Vernier / Dial / Digital (L.C.: 0.01 mm)	Using Gauge Block Set, Long Slip Gauge, Surface Plate by Comparison Method	0 to 600 mm	12 µm
191	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Surface Plate	Using Precision Level by Comparison Method	Up to 1000 mm X 1000 mm	3 x Sqrt {(L+W) / 125} µm, Where L and W are in mm
192	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Electronic Height Gauge - Linear (L.C.: 0.0001 mm)	Using Gauge Block Set, Long Slip Gauge, Surface Plate by Comparison Method	0 to 600 mm	7.5 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 104 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
193	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Electronic Height Gauge - Linear (L.C.: 0.0001 mm)	Using Caliper Checker, Surface Plate by Comparison Method	0 to 600 mm	8 µm
194	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Electronic Height Gauge - Squareness (L.C.: 0.0001 mm)	Using Granite Square Master, Digital Plunger Dial Gauge (L.C.: 0.1 µm), Surface Plate by Comparison Method	0 to 600 mm	11.8 µm
195	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile Projector - Magnification	Using Gauge Block and Digital Caliper by Comparison Method	10 X to 100 X	1 %
196	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile Projector, Video Measuring Machine - Angular (L.C.: 0.01 second of arc and Coarser)	Using Standard Angular Gauge Block By Comparison Method	0 ° to 360 °	1.9 minute of arc
197	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile Projector, Video Measuring Machine - Linear (L.C.: 0.001 mm and Coarser)	Using Linear Glass Scale by Comparison Method	0 to 200 mm	4.3 µm
198	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Tape & Scale Calibration Machine, Tape & Scale Calibrator (L.C.: 0.001 mm)	Using Gauge Block Set & Long Slip Gauge Set by Comparison Method	0 to 1000 mm	10.2 µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

105 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
199	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Barometer, Atmospheric Pressure Indicator, Atmospheric Pressure Calibrator, Manometer, Barometric Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Pump, Pressure Calibrator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	700 mbar (abs) to 1200 mbar (abs)	0.2 mbar (abs)
200	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder - Absolute Pressure	Using Pneumatic Pressure Source (Nitrogen Gas Cylinder), Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0.014 bar (abs) to 70 bar (abs)	0.05 bar (abs)
201	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Manometer - Pneumatic Absolute Pressure	Using Pneumatic Pressure Dead Weight Tester, Digital Multimeter by Direct Method as per DKD-R 6-1	0.014 bar (abs) to 30 bar (abs)	0.003 bar (abs)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 106 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
202	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Source (Nitrogen Gas Cylinder), Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 40 bar	0.007 bar
203	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 2 bar	0.0003 bar
204	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 20 bar	0.002 bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 107 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
205	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transducer, Pressure Switch, Pressure Recorder, Pressure Transmitter with Electrical Output - Pneumatic Pressure	Using Pneumatic Pressure Source (Nitrogen Gas Cylinder), Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 70 bar	0.018 bar
206	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Pressure Module with Indicator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	(-) 40 kPa to 40 kPa	0.005 kPa
207	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Pneumatic Low Pressure	Using Low Pressure Pump, Low Pressure Calibrator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 400 mbar	0.05 mbar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 108 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
208	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Hydraulic Pressure	Using Hydraulic Pressure Pump, Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 350 bar	0.047 bar
209	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Pressure Switch with Electrical Output - Hydraulic Pressure	Using Hydraulic Pressure Pump, Digital Pressure Gauge, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 700 bar	0.085 bar
210	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Vacuum Gauge, Vacuum Indicator, Vacuum Calibrator, Vacuum Controller, Vacuum Transmitter, Vacuum Transducer, Manometer, Vacuum Recorder, Vacuum Switch with Electrical Output - Pneumatic Low Vacuum	Using Low Pressure Pump, Low Pressure Calibrator, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	(-) 400 mbar to 0 mbar	0.05 mbar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 109 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
211	MECHANICAL-PRESSURE INDICATING DEVICES	Analog / Digital - Vacuum Gauge, Vacuum Indicator, Vacuum Calibrator, Vacuum Controller, Vacuum Transmitter, Vacuum Transducer, Vacuum Recorder ,Vacuum Switch with Electrical Output - Pneumatic Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	(-) 1 bar to 0 bar	0.0006 bar
212	MECHANICAL-PRESSURE INDICATING DEVICES	Analog, Digital - Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer, Pressure Recorder, Manometer - Pneumatic Absolute Pressure	Using Pneumatic Hand Pressure Pump, Multifunction Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0.014 bar (abs) to 21 bar (abs)	0.0032 bar (abs)
213	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog - Differential Pressure Gauge, Pressure Gauge, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Low Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 490 Pa	0.38 Pa



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 110 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
214	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog - Differential Pressure Gauge, Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Low Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 1034 Pa	0.35 Pa
215	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog - Differential Pressure, Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Controller, Pressure Transmitter, Pressure Transducer - Pneumatic Differential Pressure	Using Low Pressure Pump, Differential Low Pressure Calibrator, Digital Multimeter by Comparison Method as per DKD-R 6-1	0 to 2068 Pa	1.42 Pa
216	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.0001 g)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 5 kg	0.9 mg
217	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.0001 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 6 g	0.005 mg
218	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.001 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 50 g	0.005 mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 111 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
219	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 0.01 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 200 g	0.027 mg
220	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 1 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 1 kg	1.1 mg
221	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class I & Coarser (Readability : 2 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 50 kg	20 mg
222	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class II & Coarser (Readability : 10 mg)	Using E1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 6 kg	11 mg
223	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class III & Coarser (Readability : 5g)	Using F1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 75 kg	4 g
224	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class IIII (Readability : 20 g)	Using F1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 200 kg	20 g
225	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance - Class IIII (Readability : 20 g)	Using F1 Class Weights by Comparison Method as per OIML R 76-1:2006(E)	0 to 500 kg	20 g



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 112 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
226	THERMAL-SPECIFIC HEAT & HUMIDITY	Portable Temperature Indicator / Recorder / Logger	Using Standard RTD, Precision Temperature Scanner with Data Monitoring System and Temperature Chamber by Comparison Method	(-) 30 °C to 85 °C	0.13 °C
227	THERMAL-SPECIFIC HEAT & HUMIDITY	Temperature & Humidity Test / Calibration Chamber, Environment Test Chamber / Room, Climatic Test / Calibration Chamber / Room @ 23°C to 50°C - Multi Position (Minimum 9 Sensors)	Using Temperature & Humidity Recorder with Data Monitoring System by Comparison Method	10 % RH to 95 % RH	1.86 % RH
228	THERMAL-SPECIFIC HEAT & HUMIDITY	Temperature & Humidity Test / Calibration Chamber, Environment Test Chamber / Room, Climatic Test / Calibration Chamber / Room @ 25°C - Multi Position (Minimum 9 Sensors)	Using Temperature & Humidity Recorder with Data Monitoring System by Comparison Method	10 % RH to 95 % RH	1.7 % RH



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3963

Page No

113 of 119

Validity

15/08/2025 to 14/08/2029

Last Amended on

18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
229	THERMAL-SPECIFIC HEAT & HUMIDITY	Temperature & Humidity Test / Calibration Chamber, Environment Test Chamber / Room, Climatic Test / Calibration Chamber / Room @ 50%RH - Multi Position (Minimum 9 Sensors)	Using Temperature & Humidity Recorder with Data Monitoring System by Comparison Method	5 °C to 50 °C	0.85 °C
230	THERMAL-SPECIFIC HEAT & HUMIDITY	Thermo hygrometer with Internal Sensor, Digital Humidity Meter, Humidity Sensor with Indicator, Humidity Transmitter with Indicator, Temperature & Humidity Recorder @ 50% RH	Using Standard Digital RH & Temperature Indicator with Sensor, Precision Digital Multimeter, Stable Relative Humidity & Temperature Chamber by Comparison Method	5 °C to 50 °C	0.2 °C
231	THERMAL-SPECIFIC HEAT & HUMIDITY	Thermo hygrometer with Internal Sensor, Digital Humidity Meter, Temperature Humidity Sensor with Indicator, Temperature Humidity Transmitter, Temperature & Humidity Recorder @ 23°C to 50°C	Using Digital RH & Temperature Indicator with Sensor, Precision Digital Multimeter and Stable Relative Humidity & Temperature Chamber by Comparison Method	10 % RH to 95 % RH	1.88 % RH



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 114 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
232	THERMAL-SPECIFIC HEAT & HUMIDITY	Thermo hygrometer with Internal Sensor, Digital Humidity Meter, Temperature Humidity Sensor with Indicator, Temperature Humidity Transmitter, Temperature & Humidity Recorder @ 25 °C	Using Digital RH & Temperature Indicator with Sensor, Precision Digital Multimeter, Stable Relative Humidity & Temperature Chamber by Comparison Method	10 % RH to 95 % RH	0.7 % RH
233	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Dry Block Calibrator, Temperature Furnace -Single Position	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner by Comparison Method	280 °C to 660 °C	0.12 °C
234	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Dry Block Calibrator, Temperature Furnace -Single Position	Using S Type Thermocouple Sensor, Precision Temperature Scanner by Comparison Method	660 °C to 1200 °C	1.72 °C
235	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Liquid Bath Calibrator, Temperature Dry Block Calibrator - Single Position	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner by Comparison Method	140 °C to 280 °C	0.02 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 115 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
236	THERMAL-TEMPERATURE	Indicator with Sensor of Temperature Liquid Bath Calibrator, Temperature Water Bath, Temperature Dry Block Calibrator - Single Position	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner by Comparison Method	(-) 95 °C to 140 °C	0.01 °C
237	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using Non Contact Infrared Thermometer, Precision Digital Multimeter, Black Body Source by Comparison Method	> 100 °C to 500 °C	2.32 °C
238	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using Standard RTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Black Body Source by Comparison Method	(-) 20 °C to 100 °C	0.58 °C
239	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using Non Contact Infrared Thermometer, Precision Digital Multimeter, Black Body Source by Comparison Method	(-) 20 °C to 100 °C	1 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 116 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
240	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.95)	Using S - Type Thermocouple Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Black Body Source by Comparison Method	> 100 °C to 500 °C	2.12 °C
241	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.99)	Using S Type Thermocouple Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Black Body Source by Comparison Method	> 500 °C to 1200 °C	5.66 °C
242	THERMAL-TEMPERATURE	Infrared Thermometer, Infrared Temperature Sensor / Transmitter, Thermal Imaging Camera (Only Temperature) (Emissivity @ 0.99)	Using Non Contact Infrared Thermometer, Precision Digital Multimeter, Black Body Source by Comparison Method	> 500 °C to 1200 °C	7.36 °C
243	THERMAL-TEMPERATURE	Liquid-In-Glass Thermometer	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Liquid Bath by Comparison Method	(-) 80 °C to 0 °C	0.62 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 117 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
244	THERMAL-TEMPERATURE	Liquid-In-Glass Thermometer	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Liquid Bath by Comparison Method	0 °C to 150 °C	0.62 °C
245	THERMAL-TEMPERATURE	Liquid-In-Glass Thermometer	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Liquid Bath by Comparison Method	150 °C to 280 °C	0.62 °C
246	THERMAL-TEMPERATURE	Oven, Furnace - Multi Position (Minimum 9 Sensors)	Using N - Type Thermocouple with Data Monitoring System by Comparison Method	300 °C to 1200 °C	4.22 °C
247	THERMAL-TEMPERATURE	Temperature Test Chamber, Environment Test Chamber, Incubator (For Non-Medical Applications), Refrigerator, Deep Freezer, Cold Room, Storage Room - Multi Position (Minimum 9 Sensors)	Using RTD Sensor with Data Monitoring System by Comparison Method	(-) 80 °C to 180 °C	0.47 °C
248	THERMAL-TEMPERATURE	Temperature Test Chamber, Oven - Multi Position (Minimum 9 Sensors)	Using RTD Sensor with Data Monitoring System by Comparison Method	180 °C to 300 °C	0.52 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

LOORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 118 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
249	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Liquid Nitrogen bath by Comparison Method	(-) 196 °C	0.4 °C
250	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Dry Block Calibrator by Comparison Method	280 °C to 660 °C	0.16 °C
251	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Dry Block Calibrator by Comparison Method	(-) 95 °C to 140 °C	0.03 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

OORJA TECHNICAL SERVICES PVT. LTD., PLOT NO. 18, TIRUMALA
HANSREDYMADE PARK, TIGARIYA BADSHAH, INDORE INDUSTRIAL AREA,
INDORE, MADHYA PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3963

Page No 119 of 119

Validity 15/08/2025 to 14/08/2029

Last Amended on 18/08/2025

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
252	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, RTD with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using Standard PRTD Temperature Sensor, Precision Temperature Scanner, Precision Digital Multimeter and Liquid Temperature Calibrator by Comparison Method	140 °C to 280 °C	0.02 °C
253	THERMAL-TEMPERATURE	Thermocouple with / without Indicator, Temperature Gauge, Temperature Sensor with Indicator, Digital Thermometer with Sensor, Temperature Switch / Controller	Using S Type Thermocouple Sensor, Precision Temperature Scanner, Precision Digital Multimeter, Temperature Furnace Calibrator by Comparison Method	660 °C to 1200 °C	1.69 °C

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.